

REGION 2
ORGANIC PERFORMANCE EVALUATION SAMPLE
INDIVIDUAL LABORATORY SUMMARY REPORT
FOR QB 4 FY 93

LABORATORY: Industrial & Env. Analyst (NJ)
PERFORMANCE: ACCEPTABLE - No Response Required
RANK: Above = 10 Same = 3 Below = 22

% SCORE: 95.5
REPORT DATE: 09/03/93
MATRIX: WATER

COMPOUND	PREDICTION INTERVALS				LABORATORY DATA		#LABS MIS-QNT	PROGRAM #LABS NOT-ID	DATA #LABS ID-CPD	TOTAL #LABS
	WARNING LOWER	UPPER	ACTION LOWER	UPPER	CONC	Q				
TCL VOLATILE										
2-BUTANONE	130	210	120	220	140		3	0	40	40
TETRACHLOROETHENE	NU	NU	NU	NU	9		0	0	40	40
1,1,2,2-TETRACHLOROETHANE	81	110	76	120	78	\$	0	0	40	40
TOLUENE	NU	NU	NU	NU	9		0	0	40	40
STYRENE	89	110	86	110	110		2	0	40	40
TCL SEMIVOLATILE										
BIS(2-CHLOROETHYL)ETHER	19	27	18	28	23		3	0	40	40
2,2'-OXYBIS(1-CHLOROPROPANE)	18	28	17	29	30	X	3	0	40	40
NITROBENZENE	26	37	25	43	33		3	0	40	40
2,4-DIMETHYLPHENOL	32	52	29	63	32		1	0	40	40
NAPHTHALENE	16	25	15	29	22		2	0	40	40
DIMETHYL PHTHALATE	29	34	28	35	33		8	0	40	40
ACENAPHTHENE	23	31	22	32	28		3	0	40	40
2,4-DINITROPHENOL	49	82	44	87	69		6	0	40	40
4-BROMOPHENYL PHENYL ETHER	23	31	22	32	26		3	0	40	40
CHLOROBENZENE	24	32	23	36	28		4	0	40	40
ANTHRENE	16	21	16	22	18		4	0	40	40
FLUORANTHENE	18	22	17	23	18		7	0	40	40
PYRENE	48	79	43	97	64		1	0	40	40
BENZO(K)FLUORANTHENE	14	20	13	21	17		9	1	39	40
DIBENZ(A,H)ANTHRACENE	22	37	20	39	30		7	0	40	40
TCL PESTICIDES										
BETA-BHC	0.079	0.12	0.072	0.13	0.1		4	2	38	40
GAMMA-BHC (LINDANE)	0.48	0.78	0.44	0.82	0.73		4	0	40	40
ALDRIN	0.15	0.24	0.14	0.26	0.2		2	0	40	40
DIELDRIN	0.2	0.28	0.19	0.29	0.23		6	0	40	40
4,4'-DDE	0.53	0.73	0.5	0.75	0.63		3	0	40	40
ENDRIN	0.56	0.76	0.54	0.79	0.6		6	0	40	40
ENDOSULFAN SULFATE	0.53	0.75	0.5	0.78	0.6		6	1	39	40
4,4'-DDT	0.55	0.77	0.52	0.8	0.61		3	0	40	40
METHOXYCHLOR	1.2	1.8	1.1	1.9	1.3		2	0	40	40
GAMMA-CHLORDANE	0.5	0.69	0.48	0.79	0.54		2	0	40	40
NON-TCL VOLATILE										
BENZENE,DICHLORO- (O,M, OR P)					34			5	35	40
HYDRAZINE,DIMETHYL-PROPANE,1,2-DIBROMO-3-CHLORO-					0	NR		40	0	40
					18			13	27	40
NON-TCL SEMIVOLATILE										
ANILINE					0	NR		30	10	40
BENZOIC ACID					9			7	33	40
BENZYL ALCOHOL					42			3	37	40

REGION 2
ORGANIC PERFORMANCE EVALUATION SAMPLE
INDIVIDUAL LABORATORY SUMMARY REPORT
FOR QB 4 FY 93

LABORATORY: Industrial & Env. Analyst (NJ)
PERFORMANCE: ACCEPTABLE - No Response Required
RANK: Above = 10 Same = 3 Below = 22

% SCORE: 95.5
REPORT DATE: 09/03/93
MATRIX: WATER

COMPOUND	PREDICTION INTERVALS				LABORATORY		#LABS MIS-QNT	PROGRAM	DATA	TOTAL #LABS
	WARNING		ACTION		DATA			#LABS	#LABS	
	LOWER	UPPER	LOWER	UPPER	CONC	Q		NOT-ID	ID-CPD	
DIBENZOTHIOPHENE					31			0	40	40

OF TCL COMPOUNDS NOT-IDENTIFIED: 0
OF TCL COMPOUNDS MIS-QUANTIFIED: 1
OF TCL CONTAMINANTS: 0

OF NON-TCL COMPOUNDS NOT-IDENTIFIED: 0
OF NON-TCL CONTAMINANTS: 0

REGION 1
ORGANIC PERFORMANCE EVALUATION SAMPLE
INDIVIDUAL LABORATORY SUMMARY REPORT
FOR QB 4 FY 93

LABORATORY: Industrial & Env. Analyst (CT)
PERFORMANCE: ACCEPTABLE - Response Explaining Deficiency(ies) Required
RANK: Above = 26 Same = 2 Below = 7

% SCORE: 77.7
REPORT DATE: 09/03/93
MATRIX: WATER

COMPOUND	PREDICTION INTERVALS				LABORATORY DATA		#LABS MIS-QNT	PROGRAM #LABS NOT-ID	DATA #LABS ID-CPD	TOTAL #LABS
	WARNING LOWER	UPPER	ACTION LOWER	UPPER	CONC	Q				
TCL VOLATILE										
2-BUTANONE	130	210	120	220	140		3	0	40	40
TETRACHLOROETHENE	NU	NU	NU	NU	10		0	0	40	40
1,1,2,2-TETRACHLOROETHANE	81	110	76	120	98		0	0	40	40
TOLUENE	NU	NU	NU	NU	9		0	0	40	40
STYRENE	89	110	86	110	92		2	0	40	40
TCL SEMIVOLATILE										
BIS(2-CHLOROETHYL)ETHER	19	27	18	28	20		3	0	40	40
2,2'-OXYBIS(1-CHLOROPROPANE)	18	28	17	29	20		3	0	40	40
NITROBENZENE	26	37	25	43	28		3	0	40	40
2,4-DIMETHYLPHENOL	32	52	29	63	35		1	0	40	40
NAPHTHALENE	16	25	15	29	19		2	0	40	40
DIMETHYL PHTHALATE	29	34	28	35	25	X	8	0	40	40
ACENAPHTHENE	23	31	22	32	24		3	0	40	40
2,4-DINITROPHENOL	49	82	44	87	58		6	0	40	40
4-BROMOPHENYL PHENYL ETHER	23	31	22	32	23		3	0	40	40
HEXACHLOROBENZENE	24	32	23	36	22	X	4	0	40	40
PHENANTHRENE	16	21	16	22	15	X	4	0	40	40
FLUORANTHENE	18	22	17	23	15	X	7	0	40	40
PYRENE	48	79	43	97	56		1	0	40	40
BENZO(K)FLUORANTHENE	14	20	13	21	13	\$	9	1	39	40
DIBENZ(A,K)ANTHRACENE	22	37	20	39	13	X	7	0	40	40
TCL PESTICIDES										
BETA-BHC	0.079	0.12	0.072	0.13	0.091		4	2	38	40
GAMMA-BHC (LINDANE)	0.48	0.78	0.44	0.82	0.55		4	0	40	40
ALDRIN	0.15	0.24	0.14	0.26	0.18		2	0	40	40
DIELDRIN	0.2	0.28	0.19	0.29	0.21		6	0	40	40
4,4'-DDE	0.53	0.73	0.5	0.75	0.51	\$	3	0	40	40
ENDRIN	0.56	0.76	0.54	0.79	0.59		6	0	40	40
ENDOSULFAN SULFATE	0.53	0.75	0.5	0.78	0.6		6	1	39	40
4,4'-DDT	0.55	0.77	0.52	0.8	0.58		3	0	40	40
METHOXYCHLOR	1.2	1.8	1.1	1.9	1.2		2	0	40	40
GAMMA-CHLORDANE	0.5	0.69	0.48	0.79	0.55		2	0	40	40
NON-TCL VOLATILE										
BENZENE,DICHLORO- (O,M, OR P)					38			5	35	40
HYDRAZINE,DIMETHYL-					0	NR		40	0	40
PROPANE,1,2-DIBROMO-3-CHLORO-					29			13	27	40
NON-TCL SEMIVOLATILE										
ANILINE					0	NR		30	10	40
BENZOIC ACID					61			7	33	40
BENZYL ALCOHOL					36			3	37	40

REGION 1
ORGANIC PERFORMANCE EVALUATION SAMPLE
INDIVIDUAL LABORATORY SUMMARY REPORT
FOR QB 4 FY 93

LABORATORY: Industrial & Env. Analyst (CT)
PERFORMANCE: ACCEPTABLE - Response Explaining Deficiency(ies) Required
RANK: Above = 26 Same = 2 Below = 7

% SCORE: 77.7
REPORT DATE: 09/03/93
MATRIX: WATER

COMPOUND	PREDICTION INTERVALS				LABORATORY		#LABS MIS-QNT	PROGRAM #LABS NOT-ID	DATA #LABS ID-CPD	TOTAL #LABS
	WARNING		ACTION		DATA					
	LOWER	UPPER	LOWER	UPPER	CONC	Q				
DIBENZOTHIOPHENE					26			0	40	40
TCL VOLATILE (Contaminants)										
METHYLENE CHLORIDE					2			21	19	40
ACETONE					10			32	8	40
TRICHLOROETHENE					7			4	36	40
TCL SEMIVOLATILE (Contaminants)										
DIETHYLPHTHALATE					0.2			38	2	40
DI-N-BUTYLPHTHALATE					0.7			32	8	40
BIS(2-ETHYLHEXYL)PHTHALATE					0.8			30	10	40
NON-TCL SEMIVOLATILE (Contaminants)										
UNKNOWN					7			23	17	40
UNKNOWN					2			30	10	40
UNKNOWN					4			32	8	40

TCL COMPOUNDS NOT-IDENTIFIED: 0
TCL COMPOUNDS MIS-QUANTIFIED: 5
OF TCL CONTAMINANTS: 0

OF NON-TCL COMPOUNDS NOT-IDENTIFIED: 0
OF NON-TCL CONTAMINANTS: 0